



FOR OPTIMAL READING EXPERIENCE,  
PLEASE ROTATE YOUR PHONE TO LANDSCAPE MODE



INSTRUCTION FOR USE



## 1. DESCRIPTION:

Taglus Standard is a high-performance product with excellent transparency, chemical resistance, light transmission, and superior processability which makes it best suitable for the fabrication of aligner/retainer the material has been tested and approved for its biological safety.

## 2. INTENDED PURPOSE:

The Taglus Standard Thermoforming Foils can be used as aligner/retainer after thermoforming on a 3D printed model designed and planned by a registered orthodontist or registered dental lab technician.

## 3. NOTE BEFORE USING:

- Heating times may vary with current fluctuations or time.
- If the foils do not adapt well to the model, add or reduce 5 seconds to heating time until the result is ideal.
- If foils form folds, the thermoformed aligner/retainer is not clear or shows bubble formation on the thermoformed model then recalibrate your heating element or reduce the heating time until the result is ideal.

## 4. WORKING INSTRUCTIONS:

- Take 3D Printed Model required to be thermoformed. Do not remove the masking of the sheet before thermoforming.
- Place a sheet of aligner/retainer material in the pressure chamber of the thermoforming machine and enter the correct code or adjust the heating & cooling time for the Taglus sheet (see chart below- times are general guidelines only as each individual machine acts slightly different).
- Pressure should be set above 4.2 Bars and temperature is at machine default.
- Observe the sheet as it heats & as per the timer move the heating element away from the foils.
- Then close the Pressure chamber on the model for thermoforming. After completing the cooling time open the machine & see the adaptation of the sheet with a model.

WORKING INSTRUCTIONS: HEATING TIME

	0.030" (0.8mm)	0.040" (1.020mm)
<b>BioSTAR MiniSTAR/ MiniSTAR S</b>	Code 123	Code 133
<b>Dreve Druformat Scan</b>	Heating – 1:05 Cooling – 1:30	Heating – 1:10 Cooling – 1:40

- Vedia Solutions strongly recommends using thermoforming units such as the Biostar and Ministar when forming appliances with Taglus. More information on how to form Taglus appliances can be found in the Resources section of [www.taglus.com](http://www.taglus.com)
- MiniSTAR / MiniSTAR S / BioSTAR are trademarks of Scheu Dental Technology and Druformat Scan is the trademark of Dreve Dentamid.

## 5. STORAGE CONDITIONS:

- Keep the foils in the original packaging without removing the protective lamination foil during storage.
- Storage should be in a cool and dry environment at room temperature.
- Place unused sheets back in the resealable zip-lock pack along with desiccant/silica packs.

## 6. TRIMMING INSTRUCTIONS:

- First cut the excess plastic surrounding the appliance and cut Taglus appliance off the model with scissors using a trimming wheel or a twist drill. Operate the trimming wheel or twist drill at around 35,000 rpm and cut as close to the finished outline of the appliance as possible. Remove the appliance from the foil, and trim off extra parts with crown and bridge scissors or rotary burs.
- Polish the edges of the Taglus appliance with an edge polisher at around 10,000 RPM and then before delivering rinse the Taglus appliance with cool water and mild soap.

## 7. CONTRAINDICATIONS:

- This product is contraindicated for patients with a history of allergic reactions to plastic.

## 8. CLEANING:

- Products from the TAGLUS range are the best choice for daily cleaning and care. Detergents or products on the base of alcohol or plants as well as hot water or steam jets should not be used.

## 9. CAUTION:

- This product is for use by a Medical Professional, according to the instructions for use. Visually inspect and carefully check the product and packaging before use for damage during transportation and handling.
- Do not expose to extreme temperature and humidity.
- In case of any serious incident that has occurred in relation to the device please contact us at [info@taglus.com](mailto:info@taglus.com) and/or the competent authority of the Member State in which the user and/or patient is established.

## 10. SYMBOLS USED ON LABELLING:



Single Use

MD

Medical Device



Expiry Date



Caution



Batch / Lot Number



Consult Instruction for Use



Do not use if package is damage



Non-Pyrogenic



Non-Sterile



Date of Manufacture



Keep Away from Sunlight



Recycle



Keep Dry



Manufacturer



Upper & Lower Limit of Temperature



CE Mark



**Vedia Solutions**

Mfg by (A division of LDEPL)

103, Akruti Arcade, J. P. Road, Opp. A. H. Wadia School,  
Andheri West, Mumbai, Maharashtra India-400053

✉ info@taglus.com

🌐 www.taglus.com



Taglus, Taglus logo, Taglus Premium are trademarks of Vedia Solutions and its affiliates in India. Taglus logo is a registered trademark of Vedia Solutions and its affiliates in India.



EUROPE

**Wellkang Ltd**, Enterprise Hub,  
NW Business Complex,  
1 Beraghmore Rd, Derry,  
BT48 8SE, Northern Ireland